



**Sensortronic Scale Enterprises Ltd**

T-A Sensortronic Weighing & Inspection Australasia

**Client Number 7608**

PO Box 10366, Te Rapa, Hamilton, 3241  
22 Northway Street, Te Rapa, Hamilton, 3200

**Telephone 07 849-3432**

**www.swia.nz**

**Authorised Representative**

Mr Jay Attanagoda  
General Manager of Operations

**Programme**

Metrology & Calibration Laboratory

**Accreditation Number 1107**

**Initial Accreditation Date 3 July 2014**

**Conformance Standard**

ISO/IEC 17025:2017

General requirements for the competence of testing and calibration laboratories

**Laboratory Services Summary**

- 5.21 Masses
- 5.22 Precision Laboratory Balances
- 5.23 Industrial Balances
- 5.24 Industrial Weighing Appliances
- 5.51 Force Measuring Devices
- 5.53 Testing Machines

**Key Technical Personnel**

Mr Cameron Thomson	5.21, 5.22, 5.23, 5.24, 5.51, 5.53
Mr Ted Wang	5.21, 5.22, 5.23, 5.24, 5.51, 5.53

Operations Manager Authorisation:		Issue 12	Date:21/07/21	Page 1 of 4
--------------------------------------	--	----------	---------------	-------------



Sensortronic Scale Enterprises Ltd  
Metrology & Calibration Laboratory

Accreditation Number 1107

**SCOPE OF ACCREDITATION**

Calibration and Measurement Capabilities (CMC) Uncertainties are expressed as an expanded uncertainty with a level of confidence of approximately 95 % ( $k = 2$ ) <sup>Note1</sup>.

Measurement results are traceable to the International System of Units (SI) via an unbroken chain of comparisons to the New Zealand National Standards or to the National Standards of other Signatories to the CIPM MRA.

Calibrations are normally performed at the customer's premises, apart from masses.

Branch laboratories are also maintained at the following addresses:

- Unit 1, 4 Freeman Way, Manukau City, Auckland
- 17 Te Arakura Road, Feilding
- 186 Hazeldean Road, Addington, Christchurch

**5.21 Masses**

- (a) Examination of laboratory standards of mass
- (b) Examination of industrial standards of mass
- (c) Determination of the mass of solid objects

In accordance with an in-house procedure based on the Measurement Standards Laboratory of New Zealand (MSL) Technical Guide 7 and OIML R 111-1. Mass calibrations are performed at the branch laboratories in a controlled environment at  $20\text{ }^{\circ}\text{C} \pm 1\text{ }^{\circ}\text{C}$ .

Stainless steel masses	CMC Uncertainty
1 mg to 100 mg	0.02 mg to 0.05 mg
200 mg to 2 g	0.06 mg to 0.12 mg
5 g to 50 g	0.16 mg to 0.3 mg
100 g to 5 kg	$0.5 \times 10^{-5}$
10 kg to 20 kg	$1.6 \times 10^{-5}$
Cast iron masses	CMC Uncertainty
20 kg to 200 kg	$5 \times 10^{-5}$
200 kg to 500 kg	$11 \times 10^{-5}$
500 kg to 12000 kg	$7 \times 10^{-5}$

**5.22 Precision Laboratory Balances**

In accordance with an in-house procedure based on the MSL Technical Guide 25 and OIML R 111-1

Balance reading	CMC Uncertainty
1 mg to 100 mg	0.02 mg to 0.05 mg
200 mg to 2 g	0.06 mg to 0.12 mg
5 g to 50 g	0.16 mg to 0.3 mg

Operations Manager Authorisation:		Issue 12	Date:21/07/21	Page 2 of 4
--------------------------------------	--	----------	---------------	-------------



Sensortronic Scale Enterprises Ltd  
Metrology & Calibration Laboratory

Accreditation Number 1107

**SCOPE OF ACCREDITATION**

100 g to 5 kg	$0.5 \times 10^{-5}$
10 kg to 20 kg	$1.6 \times 10^{-5}$
20 kg to 65 kg	$5 \times 10^{-5}$

**5.23 Industrial Balances**

In accordance with an in-house procedure based on the MSL Technical Guide 25 and OIML R 111-1

Balance reading	CMC Uncertainty
1 mg to 100 mg	0.02 mg to 0.05 mg
200 mg to 2 g	0.06 mg to 0.12 mg
5 g to 50 g	0.16 mg to 0.3 mg
100 g to 5 kg	$0.5 \times 10^{-5}$
10 kg to 20 kg	$1.6 \times 10^{-5}$
20 kg to 200 kg	$5 \times 10^{-5}$
200 kg to 500 kg	$11 \times 10^{-5}$

**5.24 Industrial Weighing Appliances**

i) In accordance with an in-house procedure based on the MSL Technical Guide 25 and OIML R 111-1

Scale reading	CMC Uncertainty
500 kg to 12000 kg	$7 \times 10^{-5}$

ii) Dynamic weighers (checkweighers, catchweighers), in accordance with an in-house procedure based on OIML R 51-1 2006

1 mg to 100 mg	0.02 mg to 0.05 mg
200 mg to 2 g	0.06 mg to 0.12 mg
5 g to 50 g	0.16 mg to 0.3 mg
100 g to 5 kg	$0.5 \times 10^{-5}$
10 kg to 20 kg	$1.6 \times 10^{-5}$
20 kg to 100 kg	$5 \times 10^{-5}$

**5.51 Force Measuring Devices**

(b) Elastic force measuring equipment and force dynamometers

Spring balances, load cells and other mechanical or digital force gauges in accordance with an in-house procedure based on the requirements of the MSL Technical Guide 25 and OILM R 111-1.

i) By comparison with reference masses

Up to 12 tonnes (117.6 kN) in tension or compression - CMC Uncertainty as above in 5.21

Operations Manager Authorisation:		Issue 12	Date:21/07/21	Page 3 of 4
--------------------------------------	--	----------	---------------	-------------



Sensortronic Scale Enterprises Ltd  
Metrology & Calibration Laboratory

Accreditation Number 1107

**SCOPE OF ACCREDITATION**

- ii) By comparison with reference load cells  
Up to 100 tonnes (981 kN) in tension or compression - CMC Uncertainty as below in 5.53

**5.53 Testing Machines**

- (a) Tension, compression and universal machines by comparison with load cells of Class 2.0 and higher in accordance with BS EN ISO 7500-1

Load	CMC Uncertainty
49 kN	0.40 % of applied load
98 kN	0.41 % of applied load
196 kN to 610 kN	0.39 % of applied load
687 kN to 982 kN	0.38 % of applied load

**Note 1:**

Unless stated otherwise the CMC is based on the performance of the best available device and measurement uncertainties achieved for specific calibrations may be greater than the CMC Uncertainty. A laboratory may not report measurement uncertainties lower than its CMC. However, if the device under calibration has a greater accuracy than the device used to calculate the CMC the laboratory may be able to use the calibration data to lower its CMC Uncertainty. Please contact the laboratory to discuss your specific requirements.

Operations Manager Authorisation:		Issue 12	Date:21/07/21	Page 4 of 4
--------------------------------------	--	----------	---------------	-------------